



YOUR REF: 10173122
DATE RECEIVED : MAR 29, 2016
DATE REPORTED: APR 6, 2016

Report No : CTSSA/14165/16
Company :



TEST REPORT

Supplier :
Product Description : Powder Free Nitrile Examination Glove
Sample Quantity : 26 pieces
Size : Medium
Test Conducted : Force at Break During Shelf Life and After Challenge
Testing Method : EN 455 Part 2: 2009+A1:201 1
Ageing : 70 ± 2 Dec C for 7 days

SIZE	SAMPLE NO.	Force at Break, N	
		DURING SHELF LIFE	AFTER CHALLENGE
Medium	1	6.8	6.7
	2	6.3	6.1
	3	6.2	6.0
	4	6.7	6.2
	5	6.5	6.2
	6	6.3	6.0
	7	6.6	6.4
	8	6.8	6.2
	9	6.8	6.5
	10	6.4	6.1
	11	6.7	6.1
	12	6.1	6.0
	13	6.9	6.3
Median		6.3	6.2
Requirement		≥ 6.0	≥ 6.0

SGS LABORATORY SERVICES (M) SDN. BHD.

CHONG KIEN LEN
B.Sc.(HONS) AMIC
SENIOR LAB. MANAGER

Pg: 1 of 1

This document is issued by the Company under its General Conditions of Service accessible at http://www.sgs.com/terms_and_conditions.htm. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for three months only.

SGS Laboratory Services (M) Sdn. Bhd.
(Company No. 63972-M)

Consumer Testing Services No : 26, Jalan Anggerik Vanilla 31/93, Kota Kemuning, 40460 Shah Alam, Selangor Darul Ehsan.
t +6 (03) 5121 2320 f +6 (03) 5121 9082 www.sgs.com

Member of the SGS Group (SGS SA)